

湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

地址 (Add): 湖北省武汉市东湖新技术开发区茅

店山中路二号

网址 (Web site): <http://www.hbjl.gov.cn>

邮编 (Post Code): 430223

电话 (Tel): 027-81925136

传真 (Fax): 027-81925137

中国校准服务联合体

China United Calibration Service

CUC

REGISTRATION NO. 002

校准证书

CALIBRATION CERTIFICATE

证书编号: [20190902330583]

Certificate No.

委托方名称

中铁建电气化局集团南方工程有限公司

Customer

委托方地址

湖北省武汉市东湖高新区佳园路17号

Address

样品名称

直流高压发生器

Name Of Sample

制造厂商

武汉三新电力设备制造有限公司

Manufacturer

型号规格

ZGF-200kV/5mA

Model/Type

器具编号

SXB4210

No Of Sample

证书专用章

Stamp



校准日期

Calibration date

2019 年 09 月 23 日

Y M D

证书批准人

Approved by

核验员

Checked by

校准员

Calibrated by

耿睿

李亮波

田天

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



● 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

● 本次校准的技术依据 (名称、代号)

Reference documents for the Calibration (Name, Code)

参照: HJ/QC 24002-2015 直流高压发生器校准方法

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

200kV直流标准电阻分压器

数字多用表

型号/编号

Model/Serial No.

GF-200/251185

3440147M45016580

证书号/有效期

Certificate No./Due Date

(计)字

第201940043号/2021-01-13

2019002310639/2020-06-19

● 校准环境条件

Environmental condition on the Calibration

温度: 25.6℃

Temperature

气压: —

Pressure

相对湿度: 37 %

R.H.

地点: 水果湖基地高压室

Place

其它: —

Others

原始记录编号: 2019002330583

Record No.

本校准结论, 仅对受检样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.
未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

1、直流高压测量:

显示值 (kV)	实测标准值 (kV)	测量不确定度 $U_{95}, k=2$
20.0	20.01	0.3%
40.0	39.84	0.3%
60.0	59.77	0.2%
80.0	79.67	0.2%
100.0	99.54	0.2%
120.0	119.50	0.2%
140.0	139.35	0.2%
160.0	159.25	0.2%
180.0	179.13	0.2%
195.0	194.03	0.2%

2、直流电流测量 (微安表显示值):

显示值 (μA)	实测标准值 (μA)	测量不确定度 $U_{95}, k=2$
1040	1042.4	0.5%
1986	1989.5	0.5%
3591	3596.1	0.5%
4800	4805.3	0.5%

注: 建议再校期为一年。

以下空白